

附表二

導引用標準法碼最大允許誤差表

單位:mg

待檢定衡器/ 自備法碼等級 標稱值	II /F ₁	III /M ₁	III /M ₂
1000 kg	5000	50000	160000
500 kg	2500	25000	80000
200 kg	1000	10000	30000
100 kg	500	5000	16000
50 kg	250	2500	8000
20 kg	100	1000	3000
10 kg	50	500	1600
5 kg	25	250	800
2 kg	10	100	300
1 kg	5	50	160
500 g	2.5	25	80
200 g	1.0	10	30
100 g	0.5	5.0	16
50 g	0.3	3.0	10
20 g	0.25	2.5	8.0
10 g	0.20	2.0	6.0
5 g	0.16	1.6	5.0
2 g	0.12	1.2	4.0
1 g	0.10	1.0	3.0

註:表中允許誤差係指正負誤差